
## r.: 9.16.298B n: 16-04-2025

## **PROFIBUS DP**

Hook-up diagram
FLEXI-FLOW Compact Multi-channel systems
With integrated (Optional) shut-off





For reliable measurements, it is important to reduce interference from outside noise on the signals.

- When connecting to other devices, ensure that the integrity of the shielding is maintained.
- Use shielded cables and wire terminals.